


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

MAR 25 2010


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130119</u>	
Name <u>Tobin P Brennan</u>		ID# <u>4542</u>	Date <u>3-23-10</u>
A	Agency <u>Soldotna Police Department</u>		Phone # <u>(907) 262-4455</u>
Instrument Location <u>Soldotna Police Department (44510 Sterling Hwy, Soldotna Alaska 99669)</u>			
B	Alco S/N <u>X172989</u>	Target Value <u>.088</u>	High Pressure <u>1200</u>
Alco Test Values		<u>.087</u> 1 st Alco	<u>.087</u> 2 nd Alco
Signature 			
(OVER)			


BMB
4/1/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5th day of April, 2010.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

MARCH 23, 2010

OPERATOR'S NAME:

BRENNAN

OPERATOR'S NUMBER: 4542

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

UOC

O.L. #: UOC

DEPT/AGENCY: SLDI

CASE/REPORT: UOC

TEST TYPE: U

ALCO TARGET VALUE: .088

ALCO S/N: X172985

.088 ADJUSTED FOR 29.44 in
ALCO TARGET .086 19:06
BLANK TEST .000 19:07
INTERNAL STANDARD VERIFIED 19:07
ALCO TV 29.44 in .007 19:07
BLANK TEST .000 19:08
SUBJECT SAMPLE .000 19:09
BLANK TEST .000 19:09
ALCO TV 29.44 in .007 19:10
BLANK TEST .000 19:11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

MARCH 23, 2010

TIME 19:04

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 45c

BAROMETER: 29.44 in

FLOW DETECTOR: OKAY

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@AB CDEF G
HIJKLMN OPQRSTU VWXYZ[\]^_`abode fghijklmno
pqrstuvwxy z{|}~